UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.usplo.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/716,799	11/19/2003	Gary W. Guenst	P-9877.09	7018
27581 7590 04/04/2007 MEDTRONIC, INC.			EXAMINER	
710 MEDTRON	NIC PARK		GILBERT, SAMUEL G	
MINNEAPOLIS, MN 55432-9924			ART UNIT	PAPER NUMBER
			3735	
SHORTENED STATUTORY	Y PERIOD OF RESPONSE	MAIL DATE	DELIVERY MODE	
3 MON	THS	04/04/2007	PAPER	

Please find below and/or attached an Office communication concerning this application or proceeding.

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

	Application No.	Applicant(s)			
	10/716,799	GUENST ET AL.			
Office Action Summary	Examiner	Art Unit			
	Samuel G. Gilbert	3735			
The MAILING DATE of this communication app Period for Reply	ears on the cover sheet with the c	orrespondence address			
A SHORTENED STATUTORY PERIOD FOR REPLY WHICHEVER IS LONGER, FROM THE MAILING DA - Extensions of time may be available under the provisions of 37 CFR 1.13 after SIX (6) MONTHS from the mailing date of this communication. - If NO period for reply is specified above, the maximum statutory period w - Failure to reply within the set or extended period for reply will, by statute, Any reply received by the Office later than three months after the mailing earned patent term adjustment. See 37 CFR 1.704(b).	ATE OF THIS COMMUNICATION 16(a). In no event, however, may a reply be tim iii apply and will expire SIX (6) MONTHS from to cause the application to become ABANDONET	l. ely filed the mailing date of this communication. D (35 U.S.C. 6 133).			
Status .					
 Responsive to communication(s) filed on This action is FINAL. 2b) ☐ This action is non-final. Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11, 453 O.G. 213. 					
Disposition of Claims					
4) Claim(s) 1-34 is/are pending in the application. 4a) Of the above claim(s) is/are withdrawn from consideration. 5) Claim(s) is/are allowed. 6) Claim(s) 1-34 is/are rejected. 7) Claim(s) is/are objected to. 8) Claim(s) are subject to restriction and/or election requirement.					
Application Papers	•				
 9) The specification is objected to by the Examiner. 10) The drawing(s) filed on is/are: a) accepted or b) objected to by the Examiner. Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a). Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d). 11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152. 					
Priority under 35 U.S.C. § 119					
 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of: 1. Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No. 3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)). * See the attached detailed Office action for a list of the certified copies not received. 					
Attachment(s) 1) Notice of References Cited (PTO-892) 2) Notice of Draftsperson's Patent Drawing Review (PTO-948) 3) Information Disclosure Statement(s) (PTO/SB/08) Paper No(s)/Mail Date 1/20/2004.	4) Interview Summary (Paper No(s)/Mail Dat 5) Notice of Informal Pa 6) Other:	re			

DETAILED ACTION

Information Disclosure Statement

The information disclosure statement filed 1/20/2004 has been considered, the lined through references are not in English and the statement fails to comply with 37 CFR 1.98(a)(3) because it does not include a concise explanation of the relevance, as it is presently understood by the individual designated in 37 CFR 1.56(c) most knowledgeable about the content of the information, of each patent listed that is not in the English language.

One reference has not been considered because a date has not been provided.

Priority

The status of 09/879,293 should be updated in the first line of the specification.

It is unclear which application is claiming priority to the recited provisional applications because they appear to have been abandoned before the filing of this application.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Application/Control Number: 10/716,799

Art Unit: 3735

Claims 10-15, 19-23 and 27-31 are rejected under 35 U.S.C. 103(a) as being unpatentable over Spence et al (6,019,722 hereinafter Spence) in view of Hamilton et al (5,885,271 hereinafter Hamilton).

Claims 10, 19 and 27 - Spence teaches a suction pad as shown in Figure 3. The at least three outwardly extending legs are -80- and -84- which flex to the desired position, column 9 lines 41-48. Elements -82- and -86- form a flange along the legs that seals as claimed. Element -62- is a vacuum passageway. However it is not disclosed if the flange deforms during sealing, no specific material is set forth for the suction cups. Hamilton teaches silicone rubber suction cups in a heart immobilization device, column 10 lines 65 and 66. The exact material selected for suction pads is generally an obvious matter of design choice to practitioners in the medical arts. The selection of a particular material, silicone rubber of any suction pad is of no patentable significance. It would have been obvious to one of ordinary skill in the medical arts at the time the invention was made to select any known material such as silicone rubber for the suction pad of Spence as set forth in Hamilton.

Claim 11 - element -68- forms a central portion from which the legs extend.

Claims 12 and 20 - the device is insertable through a small incision between the ribs as shown in Figure 9.

Claims 13, 21, 28 and 29 - in figure 3 orifices are shown on elements -82- and -86- with ribs between some of the orifice.

Application/Control Number: 10/716,799

Art Unit: 3735

Claims 14, 15, 22, 23, 30 and 31 - a mesh is set forth in column 11 lines 5-7 of Spence. It is the examiner's position that at least metallic mesh are old and well known in the medical arts and therefore is taking official notice of such.

Double Patenting

The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. A nonstatutory obviousness-type double patenting rejection is appropriate where the conflicting claims are not identical, but at least one examined application claim is not patentably distinct from the reference claim(s) because the examined application claim is either anticipated by, or would have been obvious over, the reference claim(s). See, e.g., *In re Berg*, 140 F.3d 1428, 46 USPQ2d 1226 (Fed. Cir. 1998); *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) or 1.321(d) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent either is shown to be commonly owned with this application, or claims an invention made as a result of activities undertaken within the scope of a joint research agreement.

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

Claims 1-34 are rejected on the ground of nonstatutory obviousness-type double patenting as being unpatentable over claims 1-5, 9-12, 15-31 and 33 of U.S. Patent No. 6,676,597. Although the conflicting claims are not identical, they are not patentably distinct from each other because the differences are obvious modifications to the scope of the claims.

Art Unit: 3735

Claims 1-34 are provisionally rejected on the ground of nonstatutory obviousness-type double patenting as being unpatentable over claims 1-41 of copending Application No. 10/654,040. Although the conflicting claims are not identical, they are not patentably distinct from each other because the differences are obvious modifications in the scope of the claims.

This is a <u>provisional</u> obviousness-type double patenting rejection because the conflicting claims have not in fact been patented.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Samuel G. Gilbert whose telephone number is 571-272-4725. The examiner can normally be reached on Monday-Friday 6:30-4:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Charles Marmor II can be reached on 571-272-4730. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Application/Control Number: 10/716,799

Art Unit: 3735

Page 6

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Samuel G. Gilbert Primary Examiner Art Unit 3735